


endo★star K-FILES

1. Description:

Endostar K-files are safe instruments with four cutting edges. They are recommended for finding and shaping of narrow canals. Traditionally, this group of instruments have been made from stainless steel and they were designed in a shape proposed by Dr Kerr:

- 4 cutting edges,
- made of stainless steel,
- cross section: 
- good flexibility,
- safe, noninvasive (not too aggressive),
- ergonomic handle with ISO symbols.

2. Recommended number of use:

Maximum of 1-10 times, depending on the size (see Table 1), provided that visual inspection performed by the practitioner prior to use shows that the instrument remains undamaged, is not bent (does not apply to bending the instrument by dentist to the curvature of the canal), deformed, does not show signs of blade wear. Prolonging the life of the instrument more than recommended may result in the blade breaking.

Table 1

File No	06	08	10	15	20	25	30	35	40	45	50
Number of times that the instrument can be used	1	1	1	1-2	1-2	1-2	2-3	2-3	2-3	4-5	4-5
File No	55	60	70	80	90	100	110	120	130	140	
Number of times that the instrument can be used	4-5	4-5	4-5	4-5	8-10	8-10	8-10	8-10	8-10	8-10	

3. Cleaning and disinfection:

Detailed instructions for cleaning, disinfection and sterilization can be found on the website www.poldent.pl and www.endostar.eu in the download tab.

4. Sterilization:

This is a non-sterile product. Sterilize before use. The instruments can be sterilized in a steam sterilizer (autoclave) at 134°C. Recommended sterilization time: 3 minutes at 2.1 bar overpressure. Instruments can be disinfected with mild disinfectants and washed in ultrasonic cleaners.

5. Storage:










Instruments should be stored at room temperature in a dry, dust-free and clean environment.




6. Warnings:

This product is for professional dental use only.

7. Product claims:

Please notify the distributor and manufacturer of any claims or adverse events which occurred as a result of operating this device. Each serious incident connected with this product should be reported to the manufacturer and the competent authority of the Member State in which the user is established.

	Cross-section
	CE mark
	Sterilize in a steam autoclave at 134 ° C
	Non-sterile product
	Used for root canal preparation.
	Serial number/ Lot number
	Consult instruction for use
	Medical device
	Catalogue number

	Packaging unit
	Date of manufacture
	Stainless steel

**Manufactured by:**

Poldent Co. Ltd.

Dzika 2 Street

00-194 Warsaw, POLAND

tel.: +48 22 351-76-50 do 51

fax: +48 22 351-76-79

www.poldent.pl, www.endostar.eu, e-mail: poldent@poldent.pl, endostar@endostar.eu

Ver. 2, July 2022